

Customer	Blanken Controls B.V. Gate Dalenk Logistiek Imbosweg 30 7371 DD, Loenen, The Netherlands
Instrument	digital pressure indicator
Manufacturer	MCI
Type	M94DPV
Range	0 .. 110 kPa a // 0 .. 3000 Pa g
Serial number(s)	20071108001
Asset number	DR09
Calibration method	During the measurements, mentioned in the annexes, the device under test was directly connected to the mentioned standard. Before commencing the actual calibration a warm up time of 30 minutes was incorporated.
Reference level	pressure connection port
Location	Chrysantstraat 1 3812 WX Amersfoort, the Netherlands
Environmental conditions	Measurements were done in a climatized laboratory with the room temperature between 18°C and 23°C. The temperature variation during the calibration not exceeded ±1°C
Period of calibration	01-12-2025
Results	see page 2 (and further if applicable) of this certificate

uncertainty The reported expanded uncertainty is based on the standard uncertainty of the measurement multiplied by a coverage factor k , such that the coverage probability corresponds to approximately 95 %. The standard uncertainty has been determined in accordance with EA-4/02 M.

traceability The measurements have been executed using standards for which the traceability to (inter)national standards has been demonstrated towards the RvA.

date of issue certificate: Amersfoort, 02/12/25

RvA is member of the European Co-operation for Accreditation (EA) and is one of the signatories to the EA Multilateral Agreement and to the ILAC Mutual Recognition Arrangements (MRA) for the mutual recognition of calibration certificates.

Reproduction of the complete certificate is allowed. Parts of the certificate may only be produced with written approval of the calibration laboratory.

This certificate is issued provided that neither the Raad voor Accreditatie nor Minerva meettechniek B.V. assumes any liability.

A. Tsiporenko, corporate metrologist	type	status
	<i>digital certificate</i>	<i>original, digitally signed</i>
	<i>paper certificate</i>	<i>signed copy</i>

device under test	digital pressure indicator		
manufacturer	MCI		
type	M94DPV	channel:	1
serial number(s)	20071108001		

Description of the used pressure standard(s)

instrument	pressure controller / calibrator
manufacturer	DH Instruments
type	PPC3
platform s.n.	250
piston cylinder s.n. / range	(250-Lo) used in N2
measuring range	0 .. 207 kPa a
uncertainty of the reference pressure	10,0-10-5-p + 6 Pa

Measurement results, no adjustment performed

reference pressure [kPa a]	uncertainty of the reference pressure [kPa]	reading manual DUT [kPa a]	disagreement			expanded measurement uncertainty [kPa]
			absolute [kPa]	relative [%range]	relative [%R]	
5,0046	0,0067	5,0050	0,0004	0,000	0,008	0,0069
11,0001	0,0073	11,0000	-0,0001	0,000	-0,001	0,0074
21,9999	0,0084	22,0050	0,0051	0,005	0,023	0,0085
33,000	0,010	33,010	0,010	0,009	0,030	0,010
44,000	0,011	44,015	0,015	0,014	0,034	0,011
55,000	0,012	55,010	0,010	0,009	0,018	0,012
66,000	0,013	66,020	0,020	0,018	0,030	0,013
77,000	0,014	77,025	0,025	0,023	0,033	0,014
88,000	0,015	88,020	0,020	0,018	0,023	0,015
99,000	0,016	99,020	0,020	0,018	0,020	0,016
110,000	0,017	110,020	0,020	0,018	0,018	0,022
99,000	0,016	99,020	0,020	0,019	0,021	0,016
88,000	0,015	88,025	0,025	0,023	0,028	0,015
77,000	0,014	77,030	0,030	0,027	0,039	0,014
66,000	0,013	66,025	0,025	0,023	0,038	0,013
55,000	0,012	55,020	0,020	0,018	0,037	0,012
44,000	0,011	44,025	0,025	0,023	0,057	0,011
33,000	0,010	33,020	0,020	0,018	0,061	0,010
22,0003	0,0084	22,0150	0,0147	0,013	0,067	0,0085
10,9999	0,0073	11,0100	0,0101	0,009	0,092	0,0074
5,0034	0,0067	5,0100	0,0066	0,006	0,132	0,0069

legenda

- column 1, 2 generated reference pressure and uncertainty of the generated reference pressure
- column 3 pressure readout DUT (device under test)
- column 4 absolute difference between column 3 and column 1
- column 5,6 relative difference between column 3 and column 1
- column 7 expanded measurement uncertainty including calibration line-up, DUT resolution, repeatability, etc (k = 2)

maximum figures :	disagreement	0,027	% range
	hysteresis	0,009	% range
	uncertainty	0,020	% range

calibration technician

K.J. de Rouw

device under test	digital pressure indicator		
manufacturer	MCI		
type	M94DPV	channel: 2	
serial number(s)	20071108001		

Description of the used pressure standard(s)

instrument	pressure controller / calibrator
manufacturer	DH Instruments
type	PPC3
platform s.n.	960
piston cylinder s.n. / range	(960-Lo) used in N2
measuring range	0 .. 15 kPa g
uncertainty of the reference pressure	10,0·10 ⁻⁵ ·pe + 0,4 Pa

Measurement results, no adjustment performed

reference pressure [Pa g]	uncertainty of the reference pressure [Pa]	reading manual DUT [Pa g]	disagreement			expanded measurement uncertainty [Pa]
			absolute [Pa]	relative [%range]	relative [%R]	
-0,04	0,40	-0,04	0,00	0,000	—	0,40
300,05	0,43	299,56	-0,49	-0,016	-0,164	0,43
600,36	0,46	599,56	-0,80	-0,027	-0,133	0,46
900,10	0,49	899,46	-0,64	-0,021	-0,071	0,49
1200,14	0,52	1199,46	-0,68	-0,023	-0,057	0,52
1499,97	0,55	1499,56	-0,41	-0,014	-0,027	0,55
1800,14	0,58	1799,76	-0,38	-0,013	-0,021	0,58
2099,98	0,61	2099,46	-0,52	-0,017	-0,025	0,61
2399,90	0,64	2399,76	-0,14	-0,005	-0,006	0,64
2699,86	0,67	2699,86	0,00	0,000	0,000	0,67
2999,97	0,70	2999,56	-0,41	-0,014	-0,014	0,70
2699,84	0,67	2699,56	-0,28	-0,009	-0,010	0,67
2400,01	0,64	2399,96	-0,05	-0,002	-0,002	0,64
2100,02	0,61	2100,16	0,14	0,005	0,007	0,61
1799,98	0,58	1800,36	0,38	0,013	0,021	0,58
1500,07	0,55	1500,56	0,49	0,016	0,033	0,55
1199,87	0,52	1200,26	0,39	0,013	0,032	0,52
900,04	0,49	900,46	0,42	0,014	0,047	0,49
600,15	0,46	600,56	0,41	0,014	0,068	0,46
300,04	0,43	299,96	-0,08	-0,003	-0,027	0,43
-0,02	0,40	-0,04	-0,02	-0,001	—	0,40

legenda

column 1, 2	generated reference pressure and uncertainty of the generated reference pressure
column 3	pressure readout DUT (device under test)
column 4	absolute difference between column 3 and column 1
column 5,6	relative difference between column 3 and column 1
column 7	expanded measurement uncertainty including calibration line-up, DUT resolution, repeatability, etc (k = 2)

maximum figures :	disagreement	0,027	% range
	hysteresis	0,040	% range
	uncertainty	0,023	% range

calibration technician
K.J. de Rouw